

# CONFIRMATION

Notification: 0000040212\_05\_07  
on changes according to EN 15267 regarding certificate 0000040212\_05 dated 12 June 2019

---

**AMS designation:** Fidas 200 S, Fidas 200, Fidas 200 E for PM<sub>10</sub> and PM<sub>2.5</sub>

**Manufacturer:** PALAS GmbH  
Greschbachstrasse 3b  
76229 Karlsruhe  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 21 February 2023  
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**80 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and  
of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 44)**

The current software version of the Fidas® 200 S, Fidas® 200 E resp. Fidas® 200  
measuring systems for PM<sub>10</sub> and PM<sub>2.5</sub> from the company Palas GmbH is:


100535.0014.0001.0001.0011


The operational amplifiers on the SLA board can be installed as THT components as  
well as SMD components.

Statement issued by TÜV Rheinland Energy GmbH dated 05 September 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 25 April 2023

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.